

Research methods of hardening the surface of medical instruments

Farahov R., Kashapov R., Kashapov L., Kashapov N.
Kazan Federal University, 420008, Kremlevskaya 18, Kazan, Russia

Abstract

© Published under licence by IOP Publishing Ltd. The use of additive technologies for the manufacture of medical devices and equipment is becoming more widespread. However, manufactured products require additional processing to improve performance. Therefore, the aim was to investigate the processes of borating and electrophysical processing of the working surfaces of articles obtained by selective laser melting.

<http://dx.doi.org/10.1088/1757-899X/412/1/012088>

References

- [1] Kashapov R N, Kashapov L N and Kashapov N F 2017 Analysis and development of methods for obtaining metallic powders for selective laser melting IOP Conference Series: Materials Science and Engineering 240 012071
- [2] Kashapov R N, Kashapov L N and Kashapov N F 2015 The research of anodic microdischarges in plasma-electrolyte processing IOP Conference Series: Materials Science and Engineering 86 012019
- [3] Kashapov L N, Kashapov N F and Kashapov R N 2013 Research of the impact acidity of electrolytic cathode on the course of the plasma-electrolytic process Journal of Physics: Conference Series 479 012011
- [4] Kashapov L N, Kashapov N F and Kashapov R N 2014 Influence of plasma-electrolyte discharge to the glass surface Journal of Physics: Conference Series 567 012024
- [5] Kashapov L N, Kashapov N F, Kashapov R N and Denisov D G 2016 Plasma electrolytic treatment of products after selective laser melting Journal of Physics: Conference Series 669 012029